


HPLC COLUMN TEST SUMMARY

Catalog No:	HP-BX02-2105	Mobile Phase:	ACN:H2O=60:40
Serial No:	H22050107	Flow Rate:	0.2ml/min
Batch No:	H02020-J06-011001	Temperature:	Ambient
Bulk Material:	Hungpu XBT C18 2 μ m	Detector:	254nm
Dimensions:	50*2.1mm	Injection Volume:	1 μ l
Pressure:	2137Psi	Sample:	Acetone、Acenaphthene



Item	Sample	Ret Time (min)	USP Tailing (As)	Efficiency (N/m)
1	Acetone	0.62	1.36	40540
2	Acenaphthene	3.31	1.09	148680

Test Result: Qualified

