


HPLC COLUMN TEST SUMMARY

Catalog No:	HP-BH18-2105	Mobile Phase:	ACN:H2O=65:35
Serial No:	H22080403	Flow Rate:	0.45ml/min
Batch No:	HSS02018-J10-011001	Temperature:	Ambient
Bulk Material:	Hungpu HSS T3 1.8 μ m	Detector:	254nm
Dimensions:	50*2.1mm	Injection Volume:	1 μ l
Pressure:	5892Psi	Sample:	Acetone、Acenaphthene



Item	Sample	Ret Time (min)	USP Tailing (As)	Efficiency (N/m)
1	Acetone	0.30	1.50	64700
2	Acenaphthene	1.39	1.19	165580

Test Result:Qualified

