


HPLC COLUMN TEST SUMMARY

Catalog No:	HP-BH25-2105	Mobile Phase:	ACN:H ₂ O=70:30
Serial No:	H22080504	Flow Rate:	0.3ml/min
Batch No:	HSS02025-J10-012001	Temperature:	Ambient
Bulk Material:	Hungpu HSS T3 2.5μm	Detector:	254nm
Dimensions:	50*2.1mm	Injection Volume:	1μl
Pressure:	2246Psi	Sample:	Acetone、Acenaphthene



Item	Sample	Ret Time (min)	USP Tailing (As)	Efficiency (N/m)
1	Acetone	0.44	1.48	76620
2	Acenaphthene	1.68	1.17	163140

Test Result: Qualified

