


HPLC COLUMN TEST SUMMARY

Catalog No:	HP-BX25-0303	Mobile Phase:	ACN:H2O=50:50
Serial No:	H22072415	Flow Rate:	0.6ml/min
Batch No:	H02025-J10-011001	Temperature:	Ambient
Bulk Material:	Hungpu XBT C18 2.5 μ m	Detector:	254nm
Dimensions:	30*3.0mm	Injection Volume:	1 μ l
Pressure:	1790Psi	Sample:	Acetone、Acenaphthene



Item	Sample	Ret Time (min)	USP Tailing (As)	Efficiency (N/m)
1	Acetone	0.30	1.53	72126
2	Acenaphthene	2.77	1.06	157117

Test Result:Qualified

